

# **Conductivity Meter PCE-PWT 10**





#### Conductivity Meter PCE-PWT 10 pocket size conductivity tester / for ultrapure water / smallest measuring range 20µS/cm / salinity measurement / temperaturecompensation up to 60 °C

The conductivity meter is a good solution for anyone who wants to check the conductivity of ultrapure water. This conductivity tester is a compact device with an integrated temperature sensor for optimum temperature compensation. Among other things, the conductivity tester has a very low measuring range of only 0 ... 20  $\mu$ S / cm.

Furthermore, the cunductivity meter is able to accurately determine the salinity of liquids. The PCE-PWT 10 conductivity tester has an automatic temperature compensation in the range of 0 ... 60 ° C. The calibration of the conductivity tester is carried out via a three-point calibration.

For the desalting stage of water purification the pure water tester PCE-PWT 10 is helpfull. Separation of salt is a necessary process and can be monitored by pure water tester. The latter are a part of both, electrodialysis and reverse osmosis. Though, the purpose is the same, the principle of these process is different. Reverse osmosis is a successor of electrodialysis and if the latter uses electrical current, the former is based on pressure. For desalting of water also forward osmosis may be applied, which helps to separate the dissolved solutes from the pure water. Both methods are aimed at improving the pure water quality. This quality can be checked in the end by the pure water tester.

Conductivity is a material-specific ability to conduct electrical current. The electrolytic conductivity of aqueous solutions depends on concentration of water constituents, their degree of dissociation and the valency of the cations and anions as well as the mobility of these ions, which in turn is influenced by the temperature. Our pure water tester can accurately determine the ion content. In wastewater, the conductivity is considered as a sum parameter for the ion concentration and thus for the total salinity. With the help of our pure water tester it is possible to make a quick and easy statement about the total content of dissolved salts in the water.

- ▶ smallest measuring range up to 20 µS/cm
- compact design of pure water tester housing
- automatic temperature compensation by pure water tester
- salinity measurement
- internal three-point conductivity calibration
- changeable conductivity electrode

Subject to change




### www.pce-instruments.com

## **Specifications**

| Conductivity measuring  | 0 20 µS/cm   |
|---|--|
| range   | 0 200 µS/cm  |
|   | 0 2000 µS/cm   |
|   | 0 20.00 mS/cm  |
| Resolution  | 0.01 / 0.1 / 1   |
| Accuracy  | ± 1 % of measuring range   |
|   |  |
| Salinity measuring range  | 0.00 10.00 ppt   |
| Resolution  | 0.01 ppt   |
| Accuracy  | ± 1 % of measuring range   |
|   |  |
| Temperature measuring   | 0 60 °C / 32 140 °F  |
| range   |  |
| Resolution  | 0.1 °C   |
| Accuracy  | ± 1 °C   |
|   |  |
| Tomporature componention  |  |
| Temperature compensation  | 0 60 C / 32 140 F  |
| range   |  |
| range<br>Temperature coefficient  | 2 % / °C   |
| range<br>Temperature coefficient<br>Number of calibration points  | 2 % / °C<br>1 3  |
| range<br>Temperature coefficient  | 2 % / °C<br>1 3<br>84 μS/cm  |
| range<br>Temperature coefficient<br>Number of calibration points  | 2 % / °C<br>1 3<br>84 μS/cm<br>1413 μS/cm  |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points  | 2 % / °C<br>1 3<br>84 μS/cm<br>1413 μS/cm<br>12.88 mS/cm   |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant   | 2 % / °C<br>1 3<br>84 µS/cm<br>1413 µS/cm<br>12.88 mS/cm<br>K=1  |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant<br>Normal media temperature   | 2 % / °C<br>1 3<br>84 µS/cm<br>1413 µS/cm<br>12.88 mS/cm<br>K=1<br>25 °C / 77 °F   |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant<br>Normal media temperature<br>Hold function  | 2 % / °C<br>1 3<br>84 µS/cm<br>1413 µS/cm<br>12.88 mS/cm<br>K=1  |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant<br>Normal media temperature   | 2 % / °C<br>1 3<br>84 µS/cm<br>1413 µS/cm<br>12.88 mS/cm<br>K=1<br>25 °C / 77 °F   |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant<br>Normal media temperature<br>Hold function  | 2 % / °C<br>1 3<br>84 μS/cm<br>1413 μS/cm<br>12.88 mS/cm<br>K=1<br>25 °C / 77 °F<br>By pressing a button   |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant<br>Normal media temperature<br>Hold function<br>Automatic shutdown                          | 2 % / °C<br>1 3<br>84 μS/cm<br>1413 μS/cm<br>12.88 mS/cm<br>K=1<br>25 °C / 77 °F<br>By pressing a button<br>8 minutes after the last key press   |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant<br>Normal media temperature<br>Hold function<br>Automatic shutdown<br>Pause                 | 2 % / °C<br>1 3<br>84 µS/cm<br>1413 µS/cm<br>12.88 mS/cm<br>K=1<br>25 °C / 77 °F<br>By pressing a button<br>8 minutes after the last key press<br>Reset to factory setting possible                                      |
| range<br>Temperature coefficient<br>Number of calibration points<br>Calibration points<br>Cell constant<br>Normal media temperature<br>Hold function<br>Automatic shutdown<br>Pause<br>Power supply | 2 % / °C<br>1 3<br>84 $\mu$ S/cm<br>1413 $\mu$ S/cm<br>12.88 mS/cm<br>K=1<br>25 °C / 77 °F<br>By pressing a button<br>8 minutes after the last key press<br>Reset to factory setting possible<br>2 x 1.5 V AAA batteries |

### More information







www.pce-instruments.com